

Instruction Manual **SUNJE**

SXH-15N

(Eng)

It is recommended that only persons who have sufficient knowledge and experience such as system designers and responsible persons deal with this product after carefully reading the product manual.

1 The matters of safety

⚠ WARNING

- * Do not fabricate or do maintenance by yourself. Malfunction, Electric shock, and fire danger are considered. Contact us directly if you need the item to be repaired.
- * Do not let the water touch the product.
It may cause electric shock or fire due to malfunction.
- * When you check or maintain the product, make it sure you turn off the power.
It may cause electric shock or fire due to malfunction.
- * The item should be fixed to the target tightly.
Electric discharging or breakage might happen.
- * Please shield your installing circumstances first since the item releases the soft X-ray. Direct X-ray exposure is not recommended owing to cause health problem.
- * It includes a beryllium window on the X-ray tube unit in order to release soft X-ray for ionization. This beryllium material can cause respiratory disorder only if the person gets exposed to forms of either beryllium powder or smoke. However, our device doesn't creat beryllium in neither of the mentioned forms. Please be careful not to break the device since there might be a slight chance that the device could partially degrade into the powder form, then it can cause harm defects. If the user touches the beryllium material on the device, clean oneself on the running water or rinse it off with soap directly.
If the beryllium window gets dirty or damaged, call us directly for assistance.
- * Do not use the product at the place where dangerous material such as inflammable or ignitable material exists. This product is not the product of anti-explosive type.
- * This device is made only for industrial uses.
You need to ground the device beforehand.
Otherwise, there are in case of malfunction, electric shock, or fire damage.


⚠ CAUTION

- * Do not disconnect the head cable with power applied. There is a possibility of electric shock or malfunction.
- * Connect wires referring to the product manual.
Wrong connection can cause failures.
- * For your proper cable connection, take a look at the manual 'Installation & Connection' for references.
Any disordered connection is detected, the malfunction might arise.
- * All cables should be connected all the time. Please careful of cable disconnections.
If power or communication cables got damaged, replace them immediately.
Malfunction and fire danger may happen.
- * Do not install the device where the vibration can be detected. Any minor impact or vibration on the device, the malfunction or danger of accident might happen.
- * Do not use the item off the range of usage(Electrostatic elimination).
Malfunction or danger of accident may occur.
- * If you wish to discard the head unit of X-ray device, you need to report for disuse declaration to the institution(s) where they ask to. A proper declaration procedure is necessary in order to disuse X-ray device. Please contact to either Korea Institute of Nuclear Safety(KINS) or a safety supervisor of Sunje directly.

2 Check the package contents

▶ Product compositions

The package includes the following products composition.



* Head
SXH-15N(450)
1~4ea




* Fixing Bolt
SFB-XN-1-001
1/4"-20unc x 15L/ 1~4ea
1/4"-20unc x 8L/ 1~4ea
M4 x 0.7Pitch x 10L/ 2~8ea

▶ Option


If you require extra functions with the basic accessories we provided, you need to purchase optional items in below for further assistant.

<Changed the Direction of radiation Type>



* Head
SXH-15N(450A)


<RMS Parts>



* RMS Controller
RMSU

3 Nomenclatures of parts

▶ SXH-15N(450)



Soft X-Ray Window
The X-ray radiation will be released through this beryllium window from the head unit of X-ray ionizer.

Mounting Bracket
A bracket to fix the item on the facing target.

Controller Connector
A connector to wire the head unit and controller.

Indicator Light

RUN (Operating)

When the head unit is in operating condition, blue LED will be turned on.

TIMER (Head replacement)

When the head unit reaches the replacement cycle, The yellow LED will be turned on.


ALARM (Malfunction)

When false operation, red LED will be turned on.

▶ SXH-15N(450A)_Option

Front


Soft X-Ray Window ●
The X-ray radiation will be released through this beryllium window from the head unit of X-ray ionizer.



Mounting Bracket
A bracket to fix the item on the facing target.

Rear

Controller Connector ●
A connector to wire the head unit and controller.



Indicator Light

RUN (Operating)

When the head unit is in operating condition, blue LED will be turned on.

TIMER (Head replacement)

When the head unit reaches the replacement cycle, The yellow LED will be turned on.

ALARM (Malfunction)

When false operation, red LED will be turned on.

4 Installation and connection

▶ Check the installation conditions head unit

Please check the installation condition whether it meets proper specification.

< Must be >


- Shielded room
- There must be no objects between the target area and x-ray device.
- Installed where heat can be dissipated because operation generates the heat.
(Do not install in a confined space. It may cause malfunction)
- Be aware of any flammable or ignitable materials around the working area.
- No vibration on the fixed device.

< Advicely >

- Secure to have enough working space for quick maintenance of the device.
- The structures that thermal conduction available.
- Where the space with airflow

▶ Head unit fixing

Fix the target product on the fixed area where you install by using supported fixing bolt(s).




► Shielding methods

Shielding Material Selection

See the following graph for references.

Select the ideal tickness and quality of shielding material.



No.	Shield Material	Thickness [mm]
①	SUS	Over 0.4
②	Aluminium	Over 7.0
③	Glass	Over 10.0
④	PVC	Over 11.0
⑤	Acrylic	Over 120.0




- * Please shield the area where the customer is willing to install the soft X-ray device for safety issue.
- * After shielding process, please connection to your own interlock system to operate the X-ray device On/Off when the door is opened/closed.

► Shield installation compositions

Please refer to the "Installation Compositions" for your references.


Please keep your working environment of the leakage radiation level within the range of 10μSv/hr all the time.



► How to connect

Connecting the Head Unit

Please connect the head unit and controller with supported "Head Cable". (Max.4)




5 Problem solving

► Checklist before called


Problem	Identifying Problem	Solution
Not Operating	Is your power cable still plugged in?	Please re-plug the power cable and then turn the device on again.
	Was the proper voltage level applied for this set-up?	Please use the proper level of power input for the device. (AC 100~240V, 50/60Hz)
	Is the switch off at back of the controller?	Turn on the power switch.
	Is a fuse broken?	Replace the proper fuse immediately.
	Is there any alarm LED on the head unit of device?	Purchase and replace the head unit.
	Is remote off?	Clear the remote off corresponding to "Remote Setup" manual.
No Signal	Is Interlock off?	Place Interlock on corresponding to "Interlock setup" manual.
	Have you followed the proper procedure for Input/output signal cable(s)	Check whether the cable connection is properly done.
Decreased Performance	Can't possibly ground the device?	Please refer to "Installation & Connection" for (proper) grounding the device.
	Any damages or cut on the cables?	You need to replace the cables if it gets cut or damaged.

6 Dimensions

► SXH-15N(450)



► SXH-15N(450A)_Option



7 Specification

Parameter	Description / Value
Ion Generation Method	Soft X-ray
Source	Soft X-ray Tube
Tube Voltage	14.9kV
Beam Angle	150°
Power Consumption	Max. 15W (Based on 1 Head)
Operation Circumstance	0°C ~ +50°C (32°F ~ 122°F), 35% ~ 85% RH
Weight	Standard: 0.46kg / Option : 0.48kg
Controller	SXC-104N(450) Only
Warranty	1 year

※ The appearance and specification of the product may be changed without prior notice for the improvement of the product.

SUNJE Hi-Tek Co., Ltd.

www.sunstat.com

Head Office & Factory (Busan)
8 Cheonggwang-gil, Ilgwang-eup,
Gijang-gun, Busan, Korea
T) +82-51-720-7500 F) +82-51-720-7501

Sunje Japan Co., Ltd.
Dai 7 Matsuya Bldg 10F 1005 2-23,
Honmachi-bashi, Chuo-ku, Osaka, Japan
T) +81-6-6949-5001 F) +81-6-6949-5011

Sales Headquarter (Osan)
3rd floor, 129-20, Gyeonggi-daero 632
beon-gil, Osan-si, Gyeonggi-do, Korea
T) +82-31-203-9034 F) +82-31-202-9034

Sunje (SHANGHAI) Trading Co., Ltd.
Block C, 4F, No.482, Hongxu Road, Minhang
District, Shanghai, China
T) +86-21-5433-9761 F) +86-21-5433-9762

Customer Center +82-70-7714-9033
Sales Contact +82-31-203-9034

Sunje Technology Co., Ltd.
2F, No.6, Lane.102, Sinhe Rd, Sinfong Township,
Hsinchu County, Taiwan 30472
T) +886-3-568-7891 F) +886-3-568-7950